

W-XP WIRELESS SINGLE GAS TRANSMITTER **TOXIC, COMBUSTIBLE & VOC GASES**





TW-XP transmitters by Gastronics, Inc. sense toxic, combustible, and VOC gases in a variety of environments. TW-XP transmitters utilize sensor elements from a variety of industry leading manufacturers to ensure versatility, quality, and performance.

Sensor elements in TW-XP transmitters offer stability, longevity, response and recovery speed, resistance to fluctuations in temperature and humidity, chemical selectivity, poison resistance, linearity, repeatability, warm-up time, and alertness.

FEATURES & BENEFITS

- Durable Stainless Housing Manual Loop/Alarm Test
- Non-Intrusive Calibration
- Fast Warm Up Time
- STEL Alarm Calculation
- Independent Zero & Span
- No Factory Recharge
- Data Logging & RT Clock
- Hazardous Locations

- **Dual Relay Alarm Outputs**
- Optional 5 Relay Board
- Dual Aux Analog Inputs
- Digital Input
- IRDA Data Download
- 900 MHz, 2.4 GHz, UHF or **UHV Radio Option**

VERSATILITY

- TW-XP common board featuring various configurations
- Multiple radio options
- Auxiliary inputs
- Accepts slave board to enable up to 5 relays and 24V power supply board

DURABILITY

- All transmitters are housed in electropolished 316 stainless steel enclosures designed to withstand the harshest environments
- Housing Certifications: UL/CSA approved for Class I, Div. 1, Groups A, B, C, D and are rated NEMA-4X
- Five evenly spaced 3/4"NPT entry ports allow for a variety of installation configurations

SERVICE & DELIVERY

- Gastronics maintains a large inventory of boards, sensors and housings.
- Replacement sensors
- In the event of an emergency, packages will be hand delivered to the FedEx office (which closes at 8:00 PM).



FACTORY WARRANTY:

PART NUMBER FORMULA:

| | N/IDELESS CINICIE CAS EDANGA | | | | | | |
|---|---|--------------------------------|--|-----------------------|-----------------------|-------------|--|
| | WIRELESS SINGLE GAS TRANSM | IITTER SPECIFICA | ATIONS | | | | |
| MAIN ENCLOSURE: | 316 Stainless Steel, Electropolished, NEMA-4X, UL/CSA Class I, Div. 1, | Groups A, B, C, D | | | | | |
| ENCLOSURE DIMENSIONS: | 5.35"W x 5.25"L x 5.5"T DC Version; 7.5"T AC Version | | | | | | |
| MOUNTING OPTIONS: | Bracket Mount or Magnet Mount, Optional Deployable Tripod Availa | able | | | | | |
| DISPLAY: | 16 character by 2 line, backlit display, Operating Temp: -20C to +70C | | | | | | |
| MENU: | Non-Intrusive, Magnet Tool, Menu Driven Backlit Display. Four (4) Gr | reen/Red LED's with Hallaffect | Switch; 5 Second Gree | en LED Heartbeat & | Cold Weather Comm | nunications | |
| | Indicator | | | | | | |
| INPUT VOLTAGE: | 11-28 VDC; Option for 110/220VAC with Use of AC Board and Tall Do | | nal 3.6V "D" Size Dual | Lithum Batteries fo | r Use with Digi Radio | Version | |
| BATTERY LIFE (LITHIUM BATTERY VERSION): | Up to 26 months with Electrochemical Sensor; Up to 18 months with | | | | | | |
| SENSORS INPUT OPTION: | Voltage Output or Bridge Type Sensors; 3 Electrode Electrochemical Diffusion | Sensors | | | | | |
| SAMPLING METHOD: | Non-Intrusive with Magnet Tool | | | | | | |
| CALIBRATION: | 2 Auxiliary Analog Inputs, 1 Auxiliary Digital Input, 2 Relay Outputs | | | | | | |
| STANDARD ON BOARD I/O: SLAVE RELAY BOARD OPTION: | 5 Channel Relay Board; Designed to Operate 1 Horn and 4 Strobes w | ith On-Roard 12 to 24VDC Con | verter: Not for Use wit | h LIHE/VHE Radio D | ue to Snace Constrai | ints | |
| SLAVE RELAT BOARD OF HON. | USING SENSOR TYPE | DIGI RADIO 12V | UHF RADIO 12V | in orn / vrii Radio D | de to space constra | iiics | |
| NOMINAL CURRENT DRAW: | Toxic Gases or Oxygen or No Sensors | <35mA | <140mA | | | | |
| (detector transmitter with no slave boards or | Combustible Gases (catalytic-DD Scientific) | <55mA | <160mA | | | | |
| auxiliary I/O active) | Combustible Gases (infrared-MIPEX) | <35mA | <140mA | | | | |
| auxilial y 170 active) | VOC (Photo Ionization Detector) | <45mA | <150mA | | | | |
| | CO2 (Dynament IR) | <65mA | <170mA | | | | |
| | SENSOR SPECIFICA | | 011111 | | | | |
| | SENSOR LIST | T90 SECONDS | OP. TEMP. | OP REL. HUM. | EXPECTED LIFE | WARRANTY | |
| | Combustible Catalytic - 0-100% LEL in 1% increments | <20 | -40 to 50C | 0 to 90 NC | 24 Months | 12 | |
| | Combustible Infrared - 0-100% LEL in 1% increments | <20 | -40 to 60C | 0 to 98 NC | 60 Months | 12 | |
| | PID - 0.0-25.0 ppm in 0.1 ppm increments | <2 | -40 to 60C | 0 to 99 NC | 60 Months | 12 | |
| | Oxygen - 0.0-25.0% by volume in 0.1% increments | <15 | -40 to 60C | 15 to 90 NC | 60 Months | 24 | |
| | Ammonia - 0-100 ppm in 1 ppm increments | <60 | -40 to 40C | 15 to 90 NC | >24 Months | 12 | |
| | Carbon Monoxide - 0-300 ppm in 1 ppm increments | <30 | -40 to 50C | 15 to 90 NC | >48 Months | 24 | |
| | Hydrogen Sulfide - 0-100 ppm in 1 ppm increments | <30 | -40 to 65C | 10 to 95 NC | >48 Months | 24 | |
| | Hydrogen Sulfide - 0.0-10.0 ppm in 0.1 ppm increments | <30 | -40 to 40C | 15 to 90 NC | >48 Months | 24 | |
| STANDARD MEASURING RANGES: | Sulfur Dioxide - 0.0-20.0 ppm in 0.1 ppm increments | <25 | -20 to 50C | 15 to 90 NC | 24 Months | 12 | |
| STANDARD MEASURING NAMES. | Hydrogen Cyanide - 0.0-10.0 ppm in 0.1 ppm increments | <50 | -40 to 40C | 15 to 95 NC | >18 Months | 12 | |
| | Hydrogen Chloride - 0.0-20.0 ppm in 0.1 ppm increments | <70 | -20 to 40C | 15 to 90 NC | >24 Months | 12 | |
| | Hydrogen Fluoride - 0.0-10.0 ppm in 0.1 ppm increments | <90 | -20 to 40C | 15 to 90 NC | >18 Months | 12 | |
| | Phosgene - 0.00 - 1.00 ppm in 0.01 ppm increments | <120 | -20 to 40C | 15 to 90 NC | >12 Months | 7 | |
| | Phosphine - 0.00 - 1.00 ppm in 0.01 ppm increments | <30 | -20 to 40C | 10 to 95 NC | 24 Months | 12 | |
| | Nitrogen Dioxide - 0.0 to 20.0 ppm in 0.1 ppm increments | <30 | -20 to 40C | 15 to 90 NC | >24 Months | 12 | |
| | Nitric Oxide - 0-100 ppm in 1 ppm increments | <20 | -15 to 40C | 20 to 90 NC | >24 Months | 12 | |
| | Chlorine - 0.0-20.0 ppm in 0.1 ppm increments | <60 | -20 to 40C | 15 to 90 NC | >24 Months | 12 | |
| | Chlorine Dioxide - 0.00-5.00 in 0.01 increments | <60 | -20 to 40C | 15 to 90 NC | >18 Months | 12 | |
| | RADIO SPECIFICAT | IONS | | Diai VDaa Dua | | | |
| EDECLIENCY. | Digi XBee Pro | 2.4.647 | Digi XBee Pro 2.4 GHz | | | | |
| FREQUENCY: | 902 - 928 MHz | | | | | | |
| APPROVALS: DATA RATE: | FCC/IC (N. Am.), C-TICK (Australia), ANATEL (Brazil), IDA (Sing) | 10 Kpbs or 200 Kbps | FCC/IC (N.Am.), ETSI (EU), RCM (Australia/NZ), TELEC (Japan) | | | | |
| RF OUTPUT POWER: | 10 Kpbs or 200 Kbps Up to 250mW | · | Up to 250mW | | | | |
| SUPPLY VOLTAGE: | 2.1 to 3.6 VDC | 2.7 to 3.6 VDC | • | | | | |
| CURRENT DRAIN ON TX: | 229 Milliamps | 120 Milliamps | | | | | |
| CURRENT DRAIN ON RX: | 44 Milliamps | 31 Milliamps | | | | | |
| CURRENT DRAIN SLEEP: | 3 Microamps | 1 Microamps | · | | | | |
| RANGE: | Up to 4 Miles LOS (with High Gain Antenna) | | Up to 2 Miles LOS (with High Gain Antenna) | | | | |
| OPERATING TEMPERATURE: | -40 to 85C | -40 to 85C | -40 to 85C | | | | |
| MESHING OPTION: | Yes | Yes | | | | | |
| | Ritron DTXM UHF | | Ritron DTXM VHF | | | | |
| FREQUENCY: | 450-470MHz | 136-174MHz | | | | | |
| APPROVALS: | FCC/IC (N. American) | FCC/IC (N. American) | | | | | |
| DATA RATE: | 9600 Narrow Band, 12.5 kHz Channel Operation | | 9600 Narrow Band, 12.5 kHz Channel Operation | | | | |
| RF OUTPUT POWER: | 1-6 Watts | 1-6 Watts | | | | | |
| SUPPLY VOLTAGE: | 11-16 VDC | 11-16 VDC | | | | | |
| CURRENT DRAIN ON TX: | <1.0A at 1 Watt; <2.4A at 6 Watt | <1.0A at 1 Watt; <2.4 | A at 6 Watt | | | | |
| CURRENT DRAIN ON RX: | <120mA | <120mA | | | | | |
| CURRENT DRAIN SLEEP: | N/A | N/A | ith High Cain Auton | | | | |
| RANGE: | Up to 5 Miles LOS (with High Gain Antenna) | | Up to 10 Miles LOS (with High Gain Antenna) -30 to 50C | | | | |
| OPERATING TEMPERATURE: MESHING OPTION: | -30 to 50C | -30 to 50C No | | | | | |
| MESHING OPTION: APPROVALS: | No Class I & II, Division 1, Groups B, C, D, F, G Conforms to UL Std. 1203 | | C22 2 No. 30 & 157 T | 5 Tyne 4X | | | |
| AFFINOVALJ. | Class I, Division 1, Groups A, B, C, D Conforms to ANSI/ISA 12.12.01, U | | . CLL.Z INO. 30 Q 137, I | 5, 1ypc 4A | | | |
| | ATEX - Ex d IIB | | | | | | |
| | ALEX-EX UID II 2 G | | | | | | |
| PENDING APPROVALS: | II 2 G | | | | | | |
| | Tamb: -40°C to 65°C | | | | | | |
| | | | | | | | |
| | IFOR /C. \ IFO (CT) | | | | | | |
| | | | | | | | |
| | LISTED | | | | | | |
| FACTORY MARRANTY. | | | | | | | |

TW-WP-[Transmitter Description]-[Power Method]-[Sensor Type]-[XX]-[Flame Arrestor/Guards]-[Mounting Type]

One Year; See details on Gastronics Warranty Certificate